



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 1 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
Permanent Facility					
1	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Active Energy- 1 / 3 Phase UPF 50Hz	240V/1A to 240V/5A @ UPF , 50 Hz.	0.60%	Using 3Ph4W Energy meter by Comparison Method
2	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Active Power-1 Phase UPF to 0.2PF- 50Hz 110V to 230 V & 1A to 10A	0.1 kW to 2.3 kW	4.66% to 0.24%	Using 3Ph4W Energy meter by Comparison Method
3	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Current 100 Hz to 1kHz	0.1 mA to 1 mA	0.13% to 0.065%	Using Agilent 3458A 8 ½ DMM by Direct/Comparison method
4	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Current 100 Hz to 5kHz	1 mA to 1 A	0.065% to 0.15%	Using Agilent 3458A 8 ½ DMM by Direct/Comparison method
5	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Current 50 Hz to 100Hz	0.01 mA to 100 mA	0.53% to 0.10%	Using Agilent 3458A 8 ½ DMM by Direct/Comparison method
6	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Current 50 Hz to 100Hz	100 mA to 1 A	0.10% to 0.15%	Using Agilent 3458A 8 ½ DMM by Direct/Comparison method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 2 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
7	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Current 50 Hz to 5kHz	1 A to 10 A	0.15 % to 0.30 %	Using Fluke 8846A 6½ DMM with shunt by Direct/Comparison method
8	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Current 50 Hz to 5kHz	10 A to 20 A	0.15% to 0.65%	Using Agilent 3458A 8 ½ DMM with 30A Shunt by Direct/Comparison method
9	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC High Current - 50 Hz	20 A to 6000 A	0.35%%	Using Std CT with Fluke 8846A 6½ DMM by comparison Method
10	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC High Voltage 50Hz	1 kV to 10 kV	5.78%	Using HV Probe with DMM by comparison method
11	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Resistance	0.1 ohm to 100 kohm	0.42% to 0.15%	Using Fluke PM 6304 RCI meter by comparison Method
12	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage 50Hz to 20kHz	700 V to 1000 V	0.10%	Using Agilent 8846 A 6½ DMM by Direct/Comparison method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 3 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
13	ELECTRO- TECHNICAL- ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage 1kHz to 20 kHz	1 mV to 10 mV	0.52% to 0.09%	Using Agilent 3458A 8 ½ DMM by Direct/Comparison method
14	ELECTRO- TECHNICAL- ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage 1kHz to 20 kHz	10 mV to 500 V	0.09% to 0.075%	Using Agilent 3458A 8 ½ DMM by Direct/Comparison method
15	ELECTRO- TECHNICAL- ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage 1kHz to 20 kHz	10 mV to 10 V	0.09%	Using Agilent 3458A 8 ½ DMM by Direct/Comparison method
16	ELECTRO- TECHNICAL- ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage 20kHz to 50 kHz	1 mV to 10 mV	0.70% to 0.16%	Using Agilent 3458A 8 ½ DMM by Direct/Comparison method
17	ELECTRO- TECHNICAL- ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage 20kHz to 50 kHz	10 mV to 100 mV	0.16% to 0.09%	Using Agilent 3458A 8 ½ DMM by Direct/Comparison method
18	ELECTRO- TECHNICAL- ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage 20kHz to 50 kHz	100 mV to 100 V	0.09% to 0.045%	Using Agilent 3458A 8 ½ DMM by Direct/Comparison method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 4 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
19	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage 50 Hz to 1 kHz	1 mV to 10 mV	1.10% to 0.056%	Using Agilent 3458A 8 ½ DMM by Direct/Comparison method
20	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage 50 Hz to 1 kHz	10 mV to 10 V	0.056% to 0.012%	Using Agilent 3458A 8 ½ DMM by Direct/Comparison method
21	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage 50 Hz to 1 kHz	10 V to 700 V	0.012% to 0.05%	Using Agilent 3458A 8 ½ DMM by Direct/Comparison method
22	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	Capacitance 100 Hz	1000 nF to 100 uF	0.14% to 0.17%	Using Fluke PM 6304 RLC meter by comparison Method
23	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	Capacitance 1kHz	100 pF to 1000 nF	0.23% to 0.14%	Using Fluke PM 6304 RLC meter by comparison Method
24	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	Inductance 1KHz	100 µH to 10 H	0.38% to 0.14%	Using Fluke PM 6304 RLC Meter by comparison Method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 5 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
25	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	Power factor	0.01 PF to 1 PF	0.001PF to 0.02PF	Using oscilloscope by direct method
26	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	Power factor @50 Hz	0.2 PF to 1 PF	0.01PF	Using 3Ph/4W Energy meter by comparison method
27	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	Transformer Turn Ratio meter	0.8 ratio to 2200 ratio	0.21% to 0.25%	Using Transformer Turn Ratio Calibrator
28	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Active Power -50Hz @ 0.20 PF30V to 480V0.1A to 20A	600 mW to 1.920 kW	1.03% to 1.04%	Using Multifunction calibrator Fluke 5522A By direct method
29	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Active Power -50Hz @ 0.50 PF30V to 480V0.1A to 20A	1.50 W to 4.80 kW	0.37% to 0.39%	Using Multifunction calibrator Fluke 5522A By direct method
30	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Active Power -50Hz @ 0.80 PF30V to 480V0.1A to 20A	2.4 W to 7.680 kW	0.17% to 0.234%	Using Multifunction calibrator Fluke 5522A By direct method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 6 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
31	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Active Power -50Hz UPF30V to 480V0.001A to 0.1A	0.03 W to 3 W	0.24% to 0.08%	Using Multifunction calibrator Fluke 5522A By direct method
32	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Active Power -50Hz UPF30V to 480V0.1A to 20A	3 W to 9.60 kW	0.08% to 0.18%	Using Multifunction calibrator Fluke 5522A By direct method
33	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Capacitance -1kHz	1 µF to 600 µF	0.35%	Using High Capacitance Box By Direct Method
34	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Capacitance -1kHz	1 nF to 10 nF	1.73% to 0.41%	Using Multifunction Calibrator Fluke 5522A by Direct method
35	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Capacitance -1kHz	10 nF to 100 µF	0.40% to 0.65%	Using Multifunction calibrator Fluke 5522A By direct method
36	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Capacitance -1kHz	100 µF to 1 mF	1.53% to 1.55%	Using Multifunction calibrator Fluke 9100 By direct method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 7 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
37	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Capacitance -1kHz	100 pF to 1 μ F	0.60%	Using Discrete Capacitance Box By Direct Method
38	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Current (10 Hz to 45 Hz)	100 μ A to 3 A	0.24% to 0.21%	Using Multifunction calibrator Flue 5522A By direct method
39	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Current (10 Hz to 45 Hz)	30 μ A to 100 μ A	0.64% to 0.24%	Using Multifunction calibrator Flue 5522A By direct method
40	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Current (1kHz to 5 kHz)	1 A to 20 A	0.81% to 3.49%	Using Multifunction calibrator Flue 5522A By direct method
41	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Current (1kHz to 5 kHz)	100 mA to 1 A	0.17% to 0.81%	Using Multifunction calibrator Flue 5522A By direct method
42	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Current (1kHz to 5 kHz)	30 μ A to 100 mA	0.94% to 0.17%	Using Multifunction calibrator Flue 5522A By direct method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 8 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
43	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Current (45 Hz to 1kHz)	10 A to 20 A	0.09% to 0.16%	Using Multifunction calibrator Flue 5522A By direct method
44	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Current (45 Hz to 1kHz)	30 μ A to 10 A	0.57% to 0.09%	Using Multifunction calibrator Flue 5522A By direct method
45	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Current (50 Hz)	20 A to 1000 A	0.90% to 0.81%	Using Multifunction calibrator Flue 5522A with current coil By direct method
46	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Current (5kHz to 10 kHz)	1 A to 3 A	1.15% to 3.46%	Using Multifunction calibrator Flue 5522A By direct method
47	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Current (5kHz to 10 kHz)	100 mA to 1 A	0.35% to 1.15%	Using Multifunction calibrator Flue 5522A By direct method
48	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Current (5kHz to 10 kHz)	30 μ A to 100 mA	1.70% to 0.35%	Using Multifunction calibrator Flue 5522A By direct method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 9 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
49	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Current Harmonics	1 @ 5A to 39 @ 5A	0.77% to 2.27%	Using Multifunction calibrator Fluke 5522A By direct method @5A
50	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Inductance -1kHz	0.1 mH to 5 H	1.16% to 1.27%	Using Discrete Std Inductance Box By Direct method
51	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Power Factor 50 Hz (Lead /Lag)	0.01 PF to 1 PF	0.003PF	Using Multifunction calibrator Fluke 5522A By direct method 50Hz
52	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Resistance	100 Mohm to 1000 Mohm	0.062% to 1.82%	Using Multifunction Calibrator Fluke 5522A by Direct method
53	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Voltage (10 k Hz to 50 kHz)	3 mV to 329 V	0.39% to 0.046%	Using Multifunction calibrator Fluke 5522A By direct method
54	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Voltage (100kHz to 500kHz)	3 mV to 3 V	2.85% to 0.32%	Using Multifunction calibrator Fluke 5522A By direct method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 10 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
55	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Voltage (10Hz to 45Hz)	3 mV to 32 V	0.32% to 0.037%	Using Multifunction calibrator Flue 5522A By direct method
56	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Voltage (20kHz to 50kHz)	100 mV to 300 V	0.058% to 0.04%	Using Multifunction calibrator Flue 5522A By direct method
57	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Voltage (20kHz to 50kHz)	3 mV to 100 mV	0.42% to 0.058%	Using Multifunction calibrator Flue 5522A By direct method
58	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Voltage (45 Hz to 10kHz)	1 mV to 100 mV	0.96% to 0.028%	Using Multifunction calibrator Flue 5522A By direct method
59	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Voltage (45 Hz to 10kHz)	100 mV to 1000 V	0.028% to 0.037%	Using Multifunction calibrator Flue 5522A By direct method
60	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Voltage (50 kHz to 100 kHz)	3 mV to 329 V	0.90% to 0.27%	Using Multifunction calibrator Flue 5522A By direct method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 11 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
61	ELECTRO-TECHNICAL- ALTERNATING CURRENT (< 1 GHZ) (Source)	Voltage Harmonics @ 60 V	1 st order to 39 th order	0.77% to 2.35%	Using Multifunction calibrator Fluke 5522A By direct method
62	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	DC Current	0.01 mA to 0.1 mA	0.007% to 0.004%	Using Agilent 3458A 8 ½ DMM by Direct/Comparison method
63	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	DC Current	0.1 mA to 100 mA	0.004% to 0.005%	Using Agilent 3458A 8 ½ DMM by Direct/Comparison method
64	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	DC Current	1 A to 10 A	0.08% to 0.19%	Using Fluke 8846 6½ DMM with 30A Shunt by Direct/Comparison method
65	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	DC Current	10 A to 20 A	0.19% to 0.37%	Using Fluke 8846A 6½ DMM with Shunt by Direct/Comparison method
66	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	DC Current	100 mA to 1 A	0.005% to 0.014%	Using Agilent 3458A 8 ½ DMM by Direct/Comparison method
67	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	DC Current	20 A to 200 A	0.36% to 0.75%	Using Fluke 8846 6½ DMM with Shunt by Direct/Comparison method
68	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	DC High Voltage	1 kV to 10 kV	3.19%	Using HV probe with DMM by comparison Method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 12 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
69	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	DC High Voltage	10 kV to 20 kV	3.19%	Using HV probe with DMM by comparison Method
70	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	DC Voltage	0.1 mV to 1 mV	0.68% to 0.038%	Using Agilent 3458A 8 ½ DMM by Direct/Comparison method
71	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	DC Voltage	1 V to 1000 V	0.0012% to 0.0014%	Using Agilent 3458A 8 ½ DMM by Direct/Comparison method
72	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	DC Voltage	1 mV to 1 V	0.038% to 0.0012%	Using Agilent 3458A 8 ½ DMM by Direct/Comparison method
73	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	Resistance	1 Mohm to 100 Mohm	0.003% to 0.065%	Using Agilent 3458A 8 ½ DMM by Direct/Comparison method
74	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	Resistance	1 ohm to 10 ohm	0.099% to 0.004%	Using Agilent 3458A 8 ½ DMM by Direct/Comparison method
75	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	Resistance	10 ohm to 100 kohm	0.0029% to 0.0022%	Using Agilent 3458A 8 ½ DMM by Direct/Comparison method
76	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	Resistance	10 m ohm to 100 m ohm	1.25% to 0.66%	Using Agilent 3458A 8 ½ DMM by Direct/Comparison method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 13 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
77	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	Resistance	100 k ohm to 1 M ohm	0.0022% to 0.003%	Using Agilent 3458A 8 ½ DMM by Direct/Comparison method
78	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	Resistance	100 Mohm to 1 Gohm	0.065% to 0.6%	Using Agilent 3458A 8 ½ DMM by Direct/Comparison method
79	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	Resistance	100 mohm to 1 ohm	0.66% to 0.035%	Using Agilent 3458A 8 ½ DMM by Direct/Comparison method
80	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	DC Current	1 µA to 10 µA	2.45% to 0.25%	Using Multifunction calibrator Fluke 5522A By direct method
81	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	DC current	1 A to 10 A	0.028% to 0.064%	Using Multifunction calibrator Fluke 5522A By direct method
82	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	DC current	10 µA to 100 mA	0.25% to 0.016%	Using Multifunction calibrator Fluke 5522A By direct method
83	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	DC current	10 A to 20 A	0.064% to 0.12%	Using Multifunction calibrator Fluke 5522A By direct method
84	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	DC current	100 mA to 1 A	0.016% to 0.028%	Using Multifunction calibrator Fluke 5522A By direct method
85	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	DC current	20 A to 1000 A	0.82% to 0.71%	Using Multifunction calibrator Fluke 5522A with Current Coil By Direct Method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 14 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
86	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	DC Voltage	0.1 mV to 1 mV	1.5% to 0.24 %	Using Multifunction calibrator Fluke 5522A By direct method
87	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	DC Voltage	1 mV to 100 mV	0.24% to 0.0039%	Using Multifunction calibrator Fluke 5522A By direct method
88	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	DC Voltage	10 V to 1000 V	0.0019% to 0.0025%	Using Multifunction calibrator Fluke 5522A By direct method
89	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	DC Voltage	100 mV to 10 V	0.0039% to 0.0018%	Using Multifunction calibrator Fluke 5522A By direct method
90	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	Resistance	0.1 Mohm to 10 Mohm	0.82% to 0.58%	Using High stability Decade Megaohm Box Vaiseshika 8400HV by direct Method
91	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	Resistance	1 Gohm to 100 Gohm	1.38 % to 2.32 %	Using High stability Decade Megaohm Box 8400 HV by direct Method
92	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	Resistance	1 Mohm to 100 Mohm	0.0043% to 0.062%	Using Multifunction Calibrator Fluke 5522A by Direct method
93	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	Resistance	1 ohm to 10 kohm	0.061%	Using High Precision Decade Resistance Box Vaiseshika 7400 by direct Method
94	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	Resistance	10 Mohm to 1 Gohm	0.58% to 1.38%	Using High stability Decade Megaohm Box Vaiseshika 8400 by direct Method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 15 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
95	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	Resistance	100 Gohm to 1 Tohm	2.32% to 3.50%	Using High stability Decade Megaohm Box 8400 HV by direct Method
96	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	Resistance	100 mohm to 1 Mohm	1.52 % to 0.0043 %	Using Multifunction Calibrator Fluke 5522A by Direct method
97	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	Resistance (4wire)	10 mohm to 100 mohm	5.77% to 1.52%	Using Multifunction calibrator Fluke 5522A By Direct Method
98	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	Resistance (Micro-ohm Meters)	1 mohm @ 31.6A	0.12 %	Using Micro ohm calibrator Vaiseshika 9409-CAL by direct Method
99	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	Resistance (Micro-ohm Meters)	10 mohm @10A	0.06 %	Using Micro ohm calibrator Vaiseshika 9409-CAL by direct Method
100	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	Resistance (Micro-ohm Meters)	10μohm to 100μohm @ 200 A	4.33 % to 1.07 %	Using Micro ohm calibrator Vaiseshika 9409-CAL by direct Method
101	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	Resistance (Micro-ohm Meters)	100 mohm @3.16A	0.03 %	Using Micro ohm calibrator Vaiseshika 9409-CAL by direct Method
102	ELECTRO-TECHNICAL- ELECTRICAL EQUIPMENT (Measure)	CT Burden (1A & 5A)	1 VA to 50 VA	1.80% to 1.84%	Using Eltel Bridge(AITTS-98) by Direct Method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 16 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
103	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Measure)	CT Phase Error 50Hz	3000A/1A & 5 A to 6000A/1A & 5 A	2.25min	Using Std CT 0.05 & Eltel Bridge(AITTS-98) byComparison Method
104	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Measure)	CT Phase Error 50Hz	5A/1 & 5 A to 3200A/1 & 5 A	0.85min	Using Std CT 0.005 & Eltel Bridge(AITTS-98) byComparison Method
105	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Measure)	CT Ratio Error 50Hz	3000A /1A & 5 A to 6000A /1A & 5A A	0.084%	Using Std CT 0.05 & Eltel Bridge(AITTS-98) byComparison Method
106	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Measure)	CT Ratio Error 50Hz	5A /1 & 5A to 3200A /1 & 5A	0.015%	Using Std CT 0.005 & Eltel Bridge(AITTS-98) byComparison Method
107	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Measure)	PT Burden (63.5V & 110V)	2.5 VA to 200 VA	1.25%	Using Eltel Bridge(AITTS-98) by Direct Method
108	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Source)	Oscilloscope Bandwidth 3dB	50 kHz to 1 GHz	2.43% to 6.03%	Using Multifunction Calibrator Fluke 5522A Direct Method
109	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Source)	Oscilloscope DC Amplitude 1Mohm output	8 mV to 100 V	0.64% to 0.060%	Using Multifunction Calibrator Fluke 5522A Direct Method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 17 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
110	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Source)	Oscilloscope Time Base	1 nS to 5 S	0.076% to 0.58%	Using Multifunction Calibrator Fluke 5522A Direct Method
111	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Source)	Oscilloscope AC Amplitude 1 Mohm output @1kHz (Vpp)	8 mV to 100 V	0.98% to 0.12%	Using Multifunction Calibrator Fluke 5522A Direct Method
112	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Temperature Simulation (Indicator/Record/Controller) RTD PT-100	-200 °C to 850 °C	0.013°C to 0.026 °C	Using Agilent 3458A 8 ½ DMM using ITS-90 scale for ohm to °C conversion Method by Direct Method
113	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Temperature Simulation (Indicator/Record/Controller) B - Type	600 °C to 1820 °C	0.21°C to 0.38°C	Using Agilent 3458A 8 ½ DMM using ITS-90 scale for mV to °C conversion Method by Direct Method
114	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Temperature Simulation (Indicator/Record/Controller) C - Type	10 °C to 1820 °C	0.03°C to 0.13 °C	Using Agilent 3458A 8 ½ DMM using ITS-90 scale for mV to °C conversion Method by Direct Method
115	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Temperature Simulation (Indicator/Record/Controller) E Type	-250 °C to 1000 °C	0.124°C to 0.078°C	Using Agilent 3458A 8 ½ DMM using ITS-90 scale for mV to °C conversion Method by Direct Metho
116	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Temperature Simulation (Indicator/Record/Controller) J Type	-210 °C to 1200 °C	0.05°C to 0.10°C	Using Agilent 3458A 8 ½ DMM using ITS-90 scale for mV to °C conversion Method by Direct Metho



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 18 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
117	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Temperature Simulation (Indicator/Record/Contr oller) K Type	-200 °C to 1372 °C	0.063°C to 0.15°C	Using Agilent 3458A 8 ½ DMM using ITS-90 scale for mV to °C conversion Method by Direct Method
118	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Temperature Simulation (Indicator/Record/Contr oller) L - Type	-200 °C to 900 °C	0.02°C to 0.03°C	Using Agilent 3458A 8 ½ DMM using ITS-90 scale for mV to °C conversion Method by Direct Method
119	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Temperature Simulation (Indicator/Record/Contr oller) N Type	-200 °C to 1300 °C	0.089°C to 0.13°C	Using Agilent 3458A 8 ½ DMM using ITS-90 scale for mV to °C conversion Method by Direct Metho
120	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Temperature Simulation (Indicator/Record/Contr oller) R - Type	0 °C to 1767 °C	0.12°C to 0.36°C	Using Agilent 3458A 8 ½ DMM using ITS-90 scale for mV to °C conversion Method by Direct Method
121	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Temperature Simulation (Indicator/Record/Contr oller) S - Type	0 °C to 1767 °C	0.12°C to 0.42°C	Using Agilent 3458A 8 ½ DMM using ITS-90 scale for mV to °C conversion Method by Direct Metho
122	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Temperature Simulation (Indicator/Record/Contr oller) T- Type	-250 °C to 400 °C	0.16°C to 0.12°C	Using Agilent 3458A 8 ½ DMM using ITS-90 scale for mV to °C conversion Method by Direct Metho



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 19 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
123	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Temperature Simulation (Indicator/Record/Controller) U - Type	-200 °C to 600 °C	0.02°C to 0.06°C	Using Agilent 3458A 8 ½ DMM using ITS-90 scale for mV to °C conversion Method by Direct Method
124	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	B Type	600 °C to 1820 °C	0.21°C	Using Multifunction Calibrator Fluke 5522A Direct /using ITS-90 scale for mV to °C conversion Method
125	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	C Type	0 °C to 2316 °C	0.07°C to 0.09°C	Using Multifunction Calibrator Fluke 5522A Direct /using ITS-90 scale for mV to °C conversion Method
126	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	E Type	-250 °C to 1000 °C	0.33°C to 0.017°C	Using Multifunction Calibrator Fluke 5522A Direct /using ITS-90 scale for mV to °C conversion Method
127	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	J Type	-210 °C to 1200 °C	0.10°C to 0.09°C	Using Multifunction Calibrator Fluke 5522A Direct /using ITS-90 scale for mV to °C conversion Method
128	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	K -Type	-200 °C to 1372 °C	0.13°C to 0.09°C	Using Multifunction Calibrator Fluke 5522A Direct /using ITS-90 scale for mV to °C conversion Method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 20 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(\pm)	Remarks
129	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	L Type	-200 °C to 900 °C	0.06°C	Using Multifunction Calibrator Fluke 5522A Direct /using ITS-90 scale for mV to °C conversion Method
130	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	N Type	-200 °C to 1300 °C	0.14°C to 0.09°C	Using Multifunction Calibrator Fluke 5522A Direct /using ITS-90 scale for mV to °C conversion Method
131	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	R Type	0 °C to 1767 °C	0.31°C to 0.15°C	Using Multifunction Calibrator Fluke 5522A Direct /using ITS-90 scale for mV to °C conversion Method
132	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	RTD (Pt 100)(2 /4 wire)	-200 °C to 800 °C	0.06°C to 0.27°C	Using Multifunction Calibrator Fluke 5522A Direct /using ITS-90 scale for ohm to °C conversion Method
133	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	S Type	0 °C to 1767 °C	0.30°C to 0.17°C	Using Multifunction Calibrator Fluke 5522A Direct /using ITS-90 scale for mV to °C conversion Method
134	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	T Type	-250 °C to 400 °C	0.22°C to 0.07°C	Using Multifunction Calibrator Fluke 5522A Direct /using ITS-90 scale for mV to °C conversion Method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 21 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
135	ELECTRO-TECHNICAL- TIME & FREQUENCY (Measure)	Frequency	1 Hz to 1 GHz	0.010% to 0.006%	Using High Resolution counter by Direct method
136	ELECTRO-TECHNICAL- TIME & FREQUENCY (Measure)	Time	30 min to 24 Hrs	2.32Sec to 6Sec	Using digital Stop Watch By comparison Method
137	ELECTRO-TECHNICAL- TIME & FREQUENCY (Measure)	Time	6 sec to 30 min	0.20S to 2.32S	Using Digital Stop Watch By comparison Method
138	ELECTRO-TECHNICAL- TIME & FREQUENCY (Source)	Frequency	1 Hz to 10 Hz	0.0059% to 0.0008%	Using Multifunction Calibrator Fluke 5522A by Direct method
139	ELECTRO-TECHNICAL- TIME & FREQUENCY (Source)	Frequency	10 Hz to 1 MHz	0.0007% to 0.003%	Using Multifunction calibrator Fluke 5522A By direct method
140	ELECTRO-TECHNICAL- TIME & FREQUENCY (Source)	Frequency	100 kHz to 1 GHz	0.007% to 0.0003 %	Using Signal generator HP By Direct method
141	ELECTRO-TECHNICAL- TIME & FREQUENCY (Source)	Time & Period	2 nS to 5 S	1.17% to 0.001%	Using Multifunction calibrator Fluke 5522A By direct method
142	MECHANICAL- ACCELERATION AND SPEED	RPM Indicator / Centrifuge	100 to 5000 rpm	1 rpm to 3.5rpm	Using Digital Tachometer By comparison Method
143	MECHANICAL- ACCELERATION AND SPEED	RPM Indicator / Centrifuge	5000 to 10000 RPM	3.5 rpm to 33.0rpm	Using Digital Tachometer By comparison Method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 22 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
144	MECHANICAL-ACCELERATION AND SPEED	Tachometer , Calibrator (Non contact Type)	100 to 1000 rpm	0.2rpm to 1.9rpm	Using Digital Tachometer & RPM Calibrator By comparison Method
145	MECHANICAL-ACCELERATION AND SPEED	Tachometer , Calibrator (Non contact Type)	1000 rpm to 10000 rpm	1.9rpm to 14.8rpm	Using Digital Tachometer & RPM Calibrator By comparison Method
146	MECHANICAL-ACCELERATION AND SPEED	Tachometer , Calibrator (Non contact Type)	10000 rpm to 50000 rpm	14.74 rpm to 32.72 rpm	Using Digital Tachometer & RPM Calibrator By comparison Method
147	MECHANICAL-ACCELERATION AND SPEED	Tachometer , Calibrator (Non contact Type)	50000 to 100000 rpm	33rpm to 61.0rpm	Using Digital Tachometer & RPM Calibrator By comparison Method
148	MECHANICAL-ACCELERATION AND SPEED	Tachometer , Calibrator (contact Type)	100 rpm to 1000 rpm	1.0rpm to 2.2rpm	Using Digital Tachometer & RPM Calibrator By comparison Method
149	MECHANICAL-ACCELERATION AND SPEED	Tachometer , Calibrator (contact Type)	1000 rpm to 5000 rpm	2.2rpm to 4.0rpm	Using Digital Tachometer & RPM Calibrator By comparison Method
150	MECHANICAL-ACOUSTICS	Sound Level Meter	94 dBA & 114 dBA	2.4dBA	using Sound level Calibrator by comparison
151	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Bevel protractor LC: 5 min	0 ° to 30-45-60-90 °	4 Min	Using Angle set



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 23 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(\pm)	Remarks
152	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Bore gauge with Dial for Transmission Accuracy LC: 1 μ m	0 to 2 mm	0.9 μ m	Using ULM
153	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Caliper (Vernier/Dial/Digital)	0 to 1000 mm	19 μ m	Using Caliper checker
154	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Coating Thickness gauge	0.010 mm to 0.684 mm	12.7 μ m	Using Master foils
155	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Cylindrical Measuring pins	0.1 mm to 20 mm	0.8 μ m	Using ULM
156	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Depth Caliper (vernier/Dial/Digital) LC; 10 μ m	0 to 300 mm	9.4 μ m	Using Caliper checker ,gauge Block set, Surface Plate
157	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Depth Micrometer LC :10 μ m	0 to 300 mm	8.4 μ m	Using Slip Gauge set & Mic Check set



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 24 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
158	MECHANICAL- DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Dial Calibrator Tester LC: 0.001mm	0 to 25 mm	1.4µm	Using LVDT Probe , Slip Gauge block
159	MECHANICAL- DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Dial Gauge (Lever Type) LC: 1µm	0 to 0.8 mm	0.9µm	Using ULM
160	MECHANICAL- DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Dial Gauge (plunger Type) LC: 1µm	0 to 25 mm	2.15µm	Using Dial Calibration Tetser
161	MECHANICAL- DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Dial Gauge (plunger Type) LC: 1µm	0 mm to 10 mm	0.90 µm	Using ULM
162	MECHANICAL- DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Dial Gauge (plunger Type) LC: 10µm	0 mm to 10 mm	5.80 µm	Using ULM
163	MECHANICAL- DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Dial thickness Gauge LC: 10 µm	0 to 10 mm	7.0µm	Using slip gauge set



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 25 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
164	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Dial thickness Gauge LC; 1µm	0 to 1 mm	1.4µm	using slip gauge set
165	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	External micrometer (Inclusive of Point,Blade,Ball,Flange ,Groove, Disc, V-anvil type) LC: 1µm	0 to 100 mm	2.2µm	Using mic Check set & slip Gauge set
166	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	External Micrometer (Inclusive of Point,Blade,Ball,Flange ,Groove, Disc, V-anvil type) LC; 10µm	> 100 to 500 mm	9.9µm	Using Mic Check Set & long Slip set
167	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Feeler gauge	0.05 to 1 mm	1.3µm	Using Electronic Comparator
168	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Height gauge (Vernier/Dial/Digital) LC; 10µm	0 to 1000 mm	19µm	Using caliper Checker, Surface Plate
169	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Measuring Scale 0.5/1 mm	0 to 1000 mm	119µm	Using Tape & scale Measuring Machine



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 26 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
170	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Measuring Tape 0.5/1 mm	0 to 50000 mm	$119\sqrt{L/1000}$ µm L is in Mtr	Using Tape & scale Measuring machine
171	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Micrometer Setting Standard/ setting Rods	100 to 300 mm	5.0µm	Using Slip gauge set & LVDT Probe
172	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Micrometer Setting Standard/ setting Rods	25 to 100 mm	1.9µm	Using Slip gauge set & LVDT Probe
173	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Micrometer Setting Standard/ setting Rods	300 to 475 mm	7.8µm	Using Slip gauge set & LVDT Probe
174	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Plain Plug Gauge	100 to 300 mm	4.6µm	Using ULM
175	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Plain Plug Gauge	3 to 100 mm	1.7µm	Using ULM



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 27 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
176	MECHANICAL- DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Plain Ring Gauge	100 to 300 mm	3.4µm	Using ULM
177	MECHANICAL- DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Plain Ring Gauge	3 to 100 mm	2.7µm	Using ULM
178	MECHANICAL- DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Snap Gauge	3 to 50 mm	2.6µm	Using ULM
179	MECHANICAL- DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Snap Gauge	50 to 160 mm	3.5µm	Using ULM
180	MECHANICAL- DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Standard Foils	0 to 1.2 mm	1.3µm	Using LVDT Probe
181	MECHANICAL- DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Thread Measuring Wires	0.17 mm to 6.35 mm	1.3µm	Using Slip Gauge Set & Electronic Probe with DRO



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 28 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(\pm)	Remarks
182	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Thread Plug Gauge (Effective Diameter)	100 mm to 300 mm	4.7 μ m	Using ULM , Setting Disc , Standard Wires
183	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Thread Plug Gauge (Effective Diameter)	3 mm to 100 mm	1.3 μ m	Using ULM , Setting Disc , Standard Wires
184	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Thread Ring Gauge (Effective Diameter)	100 mm to 200 mm	3.4 μ m	Using ULM , Setting Ring
185	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Thread Ring Gauge (Effective Diameter)	200 mm to 300 mm	3.4 μ m	Using ULM , Setting Ring
186	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Thread Ring Gauge (Effective Diameter)	3 mm to 100 mm	2.7 μ m	Using ULM , Setting Ring
187	MECHANICAL-PRESSURE INDICATING DEVICES	Absolute Pressure Gauge,/Digital / Analog / Pressure Gauge, Pressure Transmitter, Pressure Switch, Barometer	0.1 to 35(Abs) bar	0.022bar	Digital Pressure Indicator with Pneumatic Pump DKD -R-6-1



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 29 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
188	MECHANICAL- PRESSURE INDICATING DEVICES	Absolute Pressure Gauge,/Digital / Analog / Pressure Gauge,Pressure Transmitter, Pressure Switch, Barometer	0.1 to 7 (Abs) bar	0.008bar	Digital Pressure Indicator with Pneumatic Pump DKD -R-6-1
189	MECHANICAL- PRESSURE INDICATING DEVICES	Absolute Pressure Gauge,/Digital / Analog / Pressure Gauge,Vacuum Gauge , Vacuum Indicator, Vacuum Transmitter, Pressure Transmitter, Pressure Switch, Barometer	60 to 110 (Abs) kPa	0.266kPa	Digital Pressure Indicator with Pneumatic Pump DKD -R-6-1
190	MECHANICAL- PRESSURE INDICATING DEVICES	Digital / Analog Pressure Gauge, Indicator ,Manometer, Magnehelic Gauge,Differential Pressure Transmitter,Pressure Switches	0 to 200(g) mbar	0.27mbar	Digital Pressure Indicator with Pneumatic Pump DKD -R-6-1
191	MECHANICAL- PRESSURE INDICATING DEVICES	Digital / Analog Pressure Gauge, Indicator ,Manometer, Magnehelic Gauge,Differential Pressure Transmitter,Pressure Switches	-12.5 mbar to 12.5(g) mbar	0.037mbar	Digital Pressure Indicator with Pneumatic Pump DKD -R-6-1



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 30 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
192	MECHANICAL-PRESSURE INDICATING DEVICES	Digital / Analog Pressure Gauge, Indicator ,Manometer, Magnehelic Gauge,Differential Pressure Transmitter,Pressure Switches	-2.5 mbar to 2.5 (g) mbar	0.014mbar	Digital Pressure Indicator with Pneumatic Pump DKD -R-6-1
193	MECHANICAL-PRESSURE INDICATING DEVICES	Digital / Analog Pressure Gauge, Indicator ,Manometer, Magnehelic Gauge,Differential Pressure Transmitter,Pressure Switches	-25 to 25(g) mbar	0.029mbar	Digital Pressure Indicator with Pneumatic Pump DKD -R-6-1
194	MECHANICAL-PRESSURE INDICATING DEVICES	Digital / Analog / Vacuum Gauge, Indicator, Vacuum Transmitter	-0.93 bar to 0 bar	0.007bar	Using Digital Vacuum gauge with Pneumatic Pump DKD-R-6-2
195	MECHANICAL-PRESSURE INDICATING DEVICES	Pressure Hydraulic Digital / Analog Pressure Gauge, Indicator, Pressure Transmitter,Pressure Switches	0 to 70 bar	0.020bar	Digital Pressure Indicator with Pneumatic Pump DKD -R-6-1
196	MECHANICAL-PRESSURE INDICATING DEVICES	Pressure HydraulicDigital / Analog Pressure Gauge, Indicator, Pressure Transmitter,Pressure Switches	0 to 350 bar	0.15bar	Digital Pressure Indicator with Hydraulic Pump DKD -R-6-1



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 31 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
197	MECHANICAL- PRESSURE INDICATING DEVICES	Pressure HydraulicDigital / Analog Pressure Gauge, Indicator, Pressure Transmitter,Pressure Switches	0 bar to 700 bar	0.19bar	Digital Pressure Indicator with Hydraulic Pump DKD -R-6-1
198	MECHANICAL- PRESSURE INDICATING DEVICES	Pressure Pneumatic Digital / Analog Pressure Gauge, Indicator, Pressure Transmitter,Pressure Switches	0 to 2 bar	0.001 bar	Digital Pressure Indicator with Pneumatic Pump DKD -R-6-1
199	MECHANICAL- PRESSURE INDICATING DEVICES	Pressure Pneumatic Digital / Analog Pressure Gauge, Indicator, Pressure Transmitter,Pressure Switches	0 to 20 bar	0.008bar	Digital Pressure Indicator with Pneumatic Pump DKD -R-6-1
200	MECHANICAL- PRESSURE INDICATING DEVICES	Pressure Pneumatic Digital / Analog Pressure Gauge, Indicator, Pressure Transmitter,Pressure Switches	0 bar to 35 bar	0.011bar	Digital Pressure Indicator with Pneumatic Pump DKD -R-6-1
201	MECHANICAL- VOLUME	Measuring Cylinder/Volumetric Flask/Graduated Jar/Can, etc.	1000 ml	0.68ml	Using standard weights of accuracy class F1 , Precision Balance (d=0.001g,d=0.01g) & distilled water of known Density by Gravimetric method based on ISO 4787



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 32 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(\pm)	Remarks
202	MECHANICAL-VOLUME	Measuring Cylinder/Volumetric Flask/Graduated Jar/Can, etc.	2000 ml	0.68ml	Using standard weights of accuracy class F1 , Precision Balance ($d=0.001g, d=0.01g$) & distilled water of known Density by Gravimetric method based on ISO 4787
203	MECHANICAL-VOLUME	Measuring Cylinder/Volumetric Flask/Graduated Jar/Can, etc.	500 ml	0.20ml	Using standard weights of accuracy class F1 , Precision Balance ($d=1\text{ mg}$) & distilled water of known Density by Gravimetric method based on ISO 4787
204	MECHANICAL-VOLUME	Micro-Pipette	1 μl to 50 μl	0.52 μl	Using Precision Balance ($d=0.001\text{mg}$ capacity 5g) & distilled water of known Density by Gravimetric method based on ISO 8655
205	MECHANICAL-VOLUME	Micro-Pipette	50 μl to 10000 μl	5 μl	Using Precision Balance ($d=0.001\text{mg}$ capacity 5g, $d=0.01\text{mg}$ capacity 210g) & distilled water of known Density by Gravimetric method based on ISO 8655



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 33 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(\pm)	Remarks
206	MECHANICAL-VOLUME	Pipette/ Burette, Measuring Cylinder /Volumetric Flask/Graduated Jar/Can,etc	1 ml	0.33 μ l	Using standard weights of accuracy class E1 , Precision Balance (d=0.001mg,d=0.01mg d=0.1mg) & distilled water of known Density by Gravimetric method based on ISO 4787
207	MECHANICAL-VOLUME	Pipette/ Burette, Measuring Cylinder /Volumetric Flask/Graduated Jar/Can,etc	10 ml	0.57 μ l	Using standard weights of accuracy class E1 , Precision Balance (d=0.001mg,d=0.01mg d=0.1mg) & distilled water of known Density by Gravimetric method based on ISO 4787
208	MECHANICAL-VOLUME	Pipette/ Burette, Measuring Cylinder /Volumetric Flask/Graduated Jar/Can,etc	100 ml	0.20ml	Using standard weights of accuracy class E1 , Precision Balance (d=1mg) & distilled water of known Density by Gravimetric method based on ISO 4787
209	MECHANICAL-VOLUME	Pipette/ Burette, Measuring Cylinder /Volumetric Flask/Graduated Jar/Can,etc	2 ml	0.33 μ l	Using standard weights of accuracy class E1 , Precision Balance (d=0.001mg,d=0.01mg d=0.1mg) & distilled water of known Density by Gravimetric method based on ISO 4787



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 34 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
210	MECHANICAL- VOLUME	Pipette/ Burette, Measuring Cylinder /Volumetric Flask/Graduated Jar/Can,etc	20 ml	1.1µl	Using standard weights of accuracy class E1 , Precision Balance (d=0.001mg,d=0.01mg d=0.1mg) & distilled water of known Density by Gravimetric method based on ISO 4787
211	MECHANICAL- VOLUME	Pipette/ Burette, Measuring Cylinder /Volumetric Flask/Graduated Jar/Can,etc	200 ml	0.2ml	Using standard weights of accuracy class E1 , Precision Balance (d=1mg) & distilled water of known Density by Gravimetric method based on ISO 4787
212	MECHANICAL- VOLUME	Pipette/ Burette, Measuring Cylinder /Volumetric Flask/Graduated Jar/Can,etc	5 ml	0.33µl	Using standard weights of accuracy class E1 , Precision Balance (d=0.001mg,d=0.01mg d=0.1mg) & distilled water of known Density by Gravimetric method based on ISO 4787
213	MECHANICAL- VOLUME	Pipette/ Burette, Measuring Cylinder /Volumetric Flask/Graduated Jar/Can,etc	50 ml	0.048ml	Using standard weights of accuracy class E1 , Precision Balance (d=1mg) & distilled water of known Density by Gravimetric method based on ISO 4787



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 35 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
214	MECHANICAL-WEIGHING SCALE AND BALANCE	Weighing Machine Readability d = 1 mg	0 kg to 20 kg	10mg	Using weights of accuracy class E1 & E 2 procedure based on OIML R-76 (2006)
215	MECHANICAL-WEIGHING SCALE AND BALANCE	Weighing Machine Readability d =0.1 mg	Upto 2300 g	0.78mg	Using weights of accuracy class E1 & E 2 procedure based on OIML R-76 (2006)
216	MECHANICAL-WEIGHING SCALE AND BALANCE	Weighing Machine Readability d =0.1g	0 kg to 35 kg	0.26g	Using weights of accuracy class E 2 procedure based on OIML R-76 (2006)
217	MECHANICAL-WEIGHING SCALE AND BALANCE	Weighing Machine Readability d=0.001mg	0 g to 5 g	0.005mg	Using weights of accuracy class E1 procedure based on OIML R-76 (2006)
218	MECHANICAL-WEIGHING SCALE AND BALANCE	Weighing Machine Readability d=0.01mg	Upto 200 g	0.06mg	Using weights of accuracy class E1 procedure based on OIML R-76 (2006)
219	MECHANICAL-WEIGHTS	Weights (Conventional Mass)	1 g	0.0032mg	Using weight of accuracy class E1, 1mg to 200g for calibration of E2 class weights & coarser with precision balance 5g/0.001mg & capacity 210g/0.01mg By substitution method ABBA weighing cycle procedure base on OIML R-111-2004



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 36 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
220	MECHANICAL-WEIGHTS	Weights (Conventional Mass)	1 kg	0.31mg	Using weight of accuracy class E2 for F1 weights & coarser with precision balance 2.3kg/0.0001g By substitution method ABBA weighing cycle procedure base on OIML R-111-2004
221	MECHANICAL-WEIGHTS	Weights (Conventional Mass)	10 g	0.011mg	Using weight of accuracy class E1, 1mg to 200g for calibration of E2 class weights & coarser with precision balance 5g/0.001mg & capacity 210g/0.01mg By substitution method ABBA weighing cycle procedure base on OIML R-111-2004
222	MECHANICAL-WEIGHTS	Weights (Conventional Mass)	10 kg	3.5mg	Using weight of accuracy class E2 for F1 weights & coarser with precision balance 23 kg/0.001g By substitution method ABBA weighing cycle procedure base on OIML R-111-2004



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 37 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(\pm)	Remarks
223	MECHANICAL-WEIGHTS	Weights (Conventional Mass)	10 mg	0.002mg	Using weight of accuracy class E1, 1mg to 200g for calibration of E2 class weights & coarser with precision balance 5g/0.001mg & capacity 210g/0.01mg By substitution method ABBA weighing cycle procedure base on OIML R-111-2004
224	MECHANICAL-WEIGHTS	Weights (Conventional Mass)	100 g	0.022mg	Using weight of accuracy class E1, 1mg to 200g for calibration of E2 class weights & coarser with precision balance 5g/0.001mg & capacity 210g/0.01mg By substitution method ABBA weighing cycle procedure base on OIML R-111-2004



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 38 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(\pm)	Remarks
225	MECHANICAL-WEIGHTS	Weights (Conventional Mass)	100 mg	0.002mg	Using weight of accuracy class E1, 1mg to 200g for calibration of E2 class weights & coarser with precision balance 5g/0.001mg & capacity 210g/0.01mg By substitution method ABBA weighing cycle procedure base on OIML R-111-2004
226	MECHANICAL-WEIGHTS	Weights (Conventional Mass)	2 g	0.004mg	Using weight of accuracy class E1, 1mg to 200g for calibration of E2 class weights & coarser with precision balance 5g/0.001mg & capacity 210g/0.01mg By substitution method ABBA weighing cycle procedure base on OIML R-111-2004
227	MECHANICAL-WEIGHTS	Weights (Conventional Mass)	2 kg	0.81mg	Using weight of accuracy class E2 for F1 weights & coarser with precision balance 2.3kg/0.0001g By substitution method ABBA weighing cycle procedure base on OIML R-111-2004



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 39 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
228	MECHANICAL-WEIGHTS	Weights (Conventional Mass)	20 g	0.015mg	Using weight of accuracy class E1, 1mg to 200g for calibration of E2 class weights & coarser with precision balance 5g/0.001mg & capacity 210g/0.01mg By substitution method ABBA weighing cycle procedure base on OIML R-111-2004
229	MECHANICAL-WEIGHTS	Weights (Conventional Mass)	20 kg	8mg	Using weight of accuracy class E2 for F1 weights & coarser with precision balance 23 kg/0.001g By substitution method ABBA weighing cycle procedure base on OIML R-111-2004
230	MECHANICAL-WEIGHTS	Weights (Conventional Mass)	20 mg	0.002mgmg	Using weight of accuracy class E1, 1mg to 200g for calibration of E2 class weights & coarser with precision balance 5g/0.001mg & capacity 210g/0.01mg By substitution method ABBA weighing cycle procedure base on OIML R-111-2004



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 40 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(\pm)	Remarks
231	MECHANICAL-WEIGHTS	Weights (Conventional Mass)	200 g	0.034mg	Using weight of accuracy class E1, 1mg to 200g for calibration of E2 class weights & coarser with precision balance 5g/0.001mg & capacity 210g/0.01mg By substitution method ABBA weighing cycle procedure base on OIML R-111-2004
232	MECHANICAL-WEIGHTS	Weights (Conventional Mass)	200 mg	0.002mg	Using weight of accuracy class E1, 1mg to 200g for calibration of E2 class weights & coarser with precision balance 5g/0.001mg & capacity 210g/0.01mg By substitution method ABBA weighing cycle procedure base on OIML R-111-2004



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 41 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(\pm)	Remarks
233	MECHANICAL-WEIGHTS	Weights (Conventional Mass)	5 g	0.005mg	Using weight of accuracy class E1, 1mg to 200g for calibration of E2 class weights & coarser with precision balance 5g/0.001mg & capacity 210g/0.01mg By substitution method ABBA weighing cycle procedure base on OIML R-111-2004
234	MECHANICAL-WEIGHTS	Weights (Conventional Mass)	5 kg	2.3mg	Using weight of accuracy class E2 for F1 weights & coarser with precision balance 23 kg/0.001g By substitution method ABBA weighing cycle procedure base on OIML R-111-2004
235	MECHANICAL-WEIGHTS	Weights (Conventional Mass)	50 g	0.022mg	Using weight of accuracy class E1, 1mg to 200g for calibration of E2 class weights & coarser with precision balance 5g/0.001mg & capacity 210g/0.01mg By substitution method ABBA weighing cycle procedure base on OIML R-111-2004



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 42 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
236	MECHANICAL-WEIGHTS	Weights (Conventional Mass)	50 mg	0.002mg	Using weight of accuracy class E1, 1mg to 200g for calibration of E2 class weights & coarser with precision balance 5g/0.001mg & capacity 210g/0.01mg By substitution method ABBA weighing cycle procedure base on OIML R-111-2004
237	MECHANICAL-WEIGHTS	Weights (Conventional Mass)	500 g	0.14mg	Using weight of accuracy class E1 for E2 weights & coarser with precision balance 2.3kg/0.0001g by substitution method ABBA weighing cycle procedure base on OIML R-111-2004
238	MECHANICAL-WEIGHTS	Weights (Conventional Mass)	500 mg	0.003mg	Using weight of accuracy class E1, 1mg to 200g for calibration of E2 class weights & coarser with precision balance 5g/0.001mg & capacity 210g/0.01mg By substitution method ABBA weighing cycle procedure base on OIML R-111-2004



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 43 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(\pm)	Remarks
239	MECHANICAL-WEIGHTS	Weights (Conventional Mass)	1 mg	0.002mg	Using weight of accuracy class E1, 1mg to 200g for calibration of E2 class weights & coarser with precision balance 5g/0.001mg & capacity 210g/0.01mg By substitution method ABBA weighing cycle procedure base on OIML R-111-2004
240	MECHANICAL-WEIGHTS	Weights (Conventional Mass)	2 mg	0.002mg	Using weight of accuracy class E1, 1mg to 200g for calibration of E2 class weights & coarser with precision balance 5g/0.001mg & capacity 210g/0.01mg By substitution method ABBA weighing cycle procedure base on OIML R-111-2004



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 44 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
241	MECHANICAL- WEIGHTS	Weights (Conventional Mass)	5 mg	0.002mg	Using weight of accuracy class E1, 1mg to 200g for calibration of E2 class weights & coarser with precision balance 5g/0.001mg & capacity 210g/0.01mg By substitution method ABBA weighing cycle procedure base on OIML R-111-2004
242	THERMAL- SPECIFIC HEAT & HUMIDITY	Humidity calibrator, Generator, environmental chamber at single point	10 to 90%RH @25°C	0.48%RH	Using RH & Temp. Probe with Indicator By Comparison Method
243	THERMAL- SPECIFIC HEAT & HUMIDITY	Humidity sensor with indicator, Transmitter with sensor, Hygrometers, Humidity data loggers with sensors	10% to 90% @25°C	0.52%RH	Using RH Generator and RH & Temp. Probe with Indicator By Comparison Method
244	THERMAL- SPECIFIC HEAT & HUMIDITY	Temperature Humidity sensor with indicator, Transmitter with sensor, ThermoHygrometers, Temperature Humidity data loggers with sensors	10 °C to 50°C @50%RH	0.35°C	Using RH & Temp. Probe with Indicator By Comparison Method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 45 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
245	THERMAL-TEMPERATURE	Data Logger, Temp.indicator with sensor, Hygrometer, Temp Transmitter with indicator	-25 °C to 28 °C	0.82°C	PRT with Precision Scanner & Negative temp chamber By Comparison Method
246	THERMAL-TEMPERATURE	Glass,Dial, DigitalThermometer	100 °C to 250 °C	0.08°C	PRT with Precision Scanner & Liquid Bath By Comparison Method
247	THERMAL-TEMPERATURE	Glass,Dial, DigitalThermometer	-80 °C to 100 °C	0.08°C	PRT with Precision Scanner & Liquid Bath By Comparison Method
248	THERMAL-TEMPERATURE	Non Contact type Infrared Thermometers , Pyrometers	0 °C to 100 °C	2.07°C	Using Std IR sensor with Indicator By comparison Method
249	THERMAL-TEMPERATURE	RTD,/PRT/TC Sensor with & Without Indicators , Digital thermometer,Temp Transmitter with Sensor, Recorder with sensor , Thermistor with temp Indicator.	(-)80 °C to 0 °C	0.046°C	PRT with Precision Scanner & Liquid Bath By Comparison Method
250	THERMAL-TEMPERATURE	RTD,/PRT/TC Sensor with & Without Indicators , Digital thermometer,Temp Transmitter with Sensor, Recorder with sensor , Thermistor with temp Indicator.	0 °C to 100 °C	0.049°C	Using PRT with Precision Scanner & Dry Bath By Comparison Method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 46 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(\pm)	Remarks
251	THERMAL-TEMPERATURE	RTD,/PRT/TC Sensor with & Without Indicators , Digital thermometer,Temp Transmitter with Sensor, Recorder with sensor ,Thermistor with temp Indicator.	0 °C to 140 °C	0.07°C	PRT with Precision Scanner & Dry Bath By Comparison Method
252	THERMAL-TEMPERATURE	RTD,/PRT/TC Sensor with & Without Indicators , Digital thermometer,Temp Transmitter with Sensor, Recorder with sensor ,Thermistor with temp Indicator.	100 °C to 200 °C	0.062°C	PRT with Precision Scanner & Liquid Bath By Comparison Method
253	THERMAL-TEMPERATURE	RTD,/PRT/TC Sensor with & Without Indicators , Digital thermometer,Temp Transmitter with Sensor, Recorder with sensor ,Thermistor with temp Indicator.	140 °C to 660 °C	0.24°C	Using PRT with Precision Scanner & Dry Bath By Comparison Method
254	THERMAL-TEMPERATURE	RTD,/PRT/TC Sensor with & Without Indicators , Digital thermometer,Temp Transmitter with Sensor, Recorder with sensor ,Thermistor with temp Indicator.	200 °C to 250 °C	0.07°C	PRT with Precision Scanner & Liquid Bath By Comparison Method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 47 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(\pm)	Remarks
255	THERMAL-TEMPERATURE	RTD,/PRT/TC Sensor with & Without Indicators , Digital thermometer,Temp Transmitter with Sensor, Recorder with sensor ,Thermistor with temp Indicator.	-95 °C to 0 °C	0.07°C	Using PRT with Precision Scanner & Dry Bath By Comparison Method
256	THERMAL-TEMPERATURE	Temp.Indicator with sensor of Deep Freezer,Oven,Chamber ,Incubators,Oil Bath,Dry well cooling Cabinet,	-95 °C to 0 °C	0.017°C to 0.042°C	Using PRT with Precision Scanner single position calibration
257	THERMAL-TEMPERATURE	Temp.Indicator with sensor of Deep Freezer,Oven,Chamber ,Incubators,Oil Bath,Dry well cooling Cabinet, Muffle Furnace .	0 °C to 660 °C	0.042°C to 0.052°C	Using PRT with Precision Scanner single position calibration
258	THERMAL-TEMPERATURE	Temp.Indicator with sensor of Deep Freezer,Oven,Chamber ,Incubators,Oil Bath,Dry well cooling Cabinet, Muffle Furnace .	1100 °C to 1200 °C	2.583°C	Using R-Type Thermocouple with Precision Scanner single position calibration



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 48 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
259	THERMAL-TEMPERATURE	Temp.Indicator with sensor of Deep Freezer,Oven,Chamber ,Incubators,Oil Bath,Dry well cooling Cabinet, Muffle Furnace .	600 °C to 800 °C	1.58°C	Using R-Type Thermocouple with Precision Scanner single position calibration
260	THERMAL-TEMPERATURE	Temp.Indicator with sensor of Deep Freezer,Oven,Chamber ,Incubators,Oil Bath,Dry well cooling Cabinet, Muffle Furnace .	800 °C to 1100 °C	1.8°C	Using R-Type Thermocouple with Precision Scanner single position calibration
261	THERMAL-TEMPERATURE	Thermocouple with & Without Indicators , Digital thermometer,Temp Transmitter with Sensor	1100 °C to 1200 °C	1.892°C to 2.583°C	Using R-Type Thermocouple with Precision scanner & furnace By Comparison Method
262	THERMAL-TEMPERATURE	Thermocouple with & Without Indicators , Digital thermometer,Temp Transmitter with Sensor	800 °C to 1100 °C	1.93°C	Using R-Type Thermocouple with Precision scanner & furnace By Comparison Method
263	THERMAL-TEMPERATURE	Thermocouple with & Without Indicators , Digital thermometer,Temp Transmitter with Sensor ,Thermistor with temp Indicator.	600 °C to 800 °C	1.77°C	Using R-Type Thermocouple with Precision scanner & furnace By Comparison Method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 49 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
Site Facility					
1	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC ACTIVE ENERGY 3PHASE /4 WIRE	240V/1A to 240V/5A @ UPF , 50 Hz.	0.94%	using Power Quality & Energy Analyser by comparison method
2	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC ACTIVE POWER	2.4 kW to 230 kW Voltage: 120 to 230VAC, Current: 100A to 1000A AC and UPF , 0.2, 0.5, 0.8 Lead , Lag	1.20%	Using Power Quality & Energy Analyser By comparison Method
3	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Current 50 Hz to 5kHz	1 A to 10 A	0.15 % to 0.30 %	Using Fluke 8846A 6½ DMM with shunt by Direct/Comparison method
4	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Current - 50 Hz to 1 kHz	10 A to 20 A	0.27% to 0.65%	Using Fluke 8846A 6½ DMM with 30A Shunt Direct method
5	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Current - 50 Hz to 1 kHz	100 µA to 10 A	0.09% to 0.19%	Using Fluke 8846A 6½ DMM Direct method
6	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Current - 50 Hz to 1 kHz	30 µA to 100 µA	0.62% to 0.16%	Using Fluke 8846A 6½ DMM Direct method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 50 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
Site Facility					
7	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC High Current - 50 Hz	20 A to 6000 A	0.35%%	Using Std CT with Fluke 8846A 6½ DMM by comparison Method
8	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC High Voltage 50Hz	1 kV to 10 kV	5.78%	Using HV probe with DMM by comparison Method
9	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC High Voltage 50Hz	10 kV to 100 kV	3.57%	Using HV Divider with DMM by comparison Method
10	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage 50Hz to 1KHz	0.1 mV to 1 mV	4.75% to 0.42%	Using Fluke 8846A 6½ DMM Direct method
11	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage 50Hz to 1KHz	1 mV to 10 mV	0.42% to 0.53%	Using Fluke 8846A 6½ DMM Direct method
12	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage 50Hz to 1KHz	10 mV to 10 V	0.54% to 0.12%	Using Fluke 8846A 6½ DMM Direct method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 51 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
Site Facility					
13	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage 50Hz to 1KHz	10 V to 1000 V	0.12% to 0.10%	Using Fluke 8846A 6½ DMM Direct method
14	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage @ 1kHz to 50 kHz	50 mV to 100 V	0.27% to 0.33%	Using Fluke 8846A 6½ DMM Direct method
15	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	Capacitance 100 Hz	100 µF to 10 mF	0.13%	Using RCL meter Fluke PM6304 by direct method
16	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	Capacitance 1kHz	100 pF to 100 µF	0.23% to 0.12%	Using RCL meter Fluke PM6304 by direct method
17	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	Impulse Voltage (-ve & +ve Peak)	1.2 V to 100 V and 1.2µsec to 100µsec	2.80%	Using RF HV probe with Oscilloscope by comparison Method
18	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	Impulse Voltage (-ve & +ve Peak)	100 V to 10 kV	2.80%	Using RF HV probe with Oscilloscope by comparison Method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 52 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
Site Facility					
19	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	Inductance 1kHz	100 µH to 10 H	0.38% to 0.14%	Using RLC Fluke PM6304 meter by Direct /comparison Method
20	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	Power factor	0.01 PF to 1 PF	0.001PF to 0.02PF	Using oscilloscope by direct method
21	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	POWER FACTOR	UPF to 0.20 PF	0.035PF	Using Power Quality & Energy Analyser By comparison Method
22	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	Resistance AC	0.1 ohm to 100 kohm	0.12% to 0.15%	Using Fluke PM 6304 RCL meter by comparison Method
23	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	Transformer Turn Ratio meter	0.8 ratio to 2200 ratio	0.21% to 0.25%	Using Transformer Turn Ratio Calibrator
24	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Active Power -50Hz @ 0.20 PF30V to 480V0.1A to 20A	600 mW to 1.920 kW	1.03% to 1.04%	Using Multifunction calibrator Fluke 5522A By direct method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 53 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
Site Facility					
25	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Active Power -50Hz @ 0.50 PF30V to 480V0.1A to 20A	1.50 W to 4.80 kW	0.37% to 0.39%	Using Multifunction calibrator Fluke 5522A By direct method
26	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Active Power -50Hz @ 0.80 PF30V to 480V0.1A to 20A	2.4 W to 7.680 kW	0.17% to 0.234%	Using Multifunction calibrator Fluke 5522A By direct method
27	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Active Power -50Hz UPF30V to 480V0.001A to 0.1A	0.03 W to 3 W	0.24% to 0.08%	Using Multifunction calibrator Fluke 5522A By direct method
28	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Active Power -50Hz UPF30V to 480V0.1A to 20A	3 W to 9.60 kW	0.08% to 0.18%	Using Multifunction calibrator Fluke 5522A By direct method
29	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Capacitance -1kHz	1 µF to 600 µF	0.35%	Using High Capacitance Box By Direct Method
30	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Capacitance -1kHz	1 nF to 10 nF	1.73% to 0.41%	Using Multifunction Calibrator Fluke 5522A by Direct method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 54 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
Site Facility					
31	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Capacitance -1kHz	10 nF to 100 μ F	0.40% to 0.65%	Using Multifunction calibrator Fluke 5522A By direct method
32	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Capacitance -1kHz	100 μ F to 1 mF	1.53%% to 1.55%	Using Multifunction calibrator Fluke 9100 By direct method
33	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Capacitance -1kHz	100 pF to 1 μ F	0.60%	Using Discrete Capacitance Box By Direct Method
34	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Current (10 Hz to 45 Hz)	100 μ A to 3 A	0.24% to 0.21%	Using Multifunction calibrator Flue 5522A By direct method
35	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Current (10 Hz to 45 Hz)	30 μ A to 100 μ A	0.64% to 0.24%	Using Multifunction calibrator Flue 5522A By direct method
36	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Current (1kHz to 5 kHz)	1 A to 20 A	0.81% to 3.49%	Using Multifunction calibrator Flue 5522A By direct method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 55 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
Site Facility					
37	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Current (1kHz to 5 kHz)	100 mA to 1 A	0.17% to 0.81%	Using Multifunction calibrator Flue 5522A By direct method
38	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Current (1kHz to 5 kHz)	30 μ A to 100 mA	0.94% to 0.17%	Using Multifunction calibrator Flue 5522A By direct method
39	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Current (45 Hz to 1kHz)	10 A to 20 A	0.09% to 0.16%	Using Multifunction calibrator Flue 5522A By direct method
40	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Current (45 Hz to 1kHz)	30 μ A to 10 A	0.57% to 0.09%	Using Multifunction calibrator Flue 5522A By direct method
41	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Current (50 Hz)	20 A to 1000 A	0.90% to 0.81%	Using Multifunction calibrator Flue 5522A with current coil By direct method
42	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Current (5kHz to 10 kHz)	1 A to 3 A	1.15% to 3.46%	Using Multifunction calibrator Flue 5522A By direct method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 56 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
Site Facility					
43	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Current (5kHz to 10 kHz)	100 mA to 1 A	0.35% to 1.15%	Using Multifunction calibrator Fluke 5522A By direct method
44	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Current (5kHz to 10 kHz)	30 μ A to 100 mA	1.70% to 0.35%	Using Multifunction calibrator Fluke 5522A By direct method
45	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Current Harmonics	1 @ 5A to 39 @ 5A	0.77% to 2.27%	Using Multifunction calibrator Fluke 5522A By direct method @5A
46	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Inductance -1kHz	0.1 mH to 5 H	1.16% to 1.27%	Using Discrete Std Inductance Box By Direct method
47	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Power Factor 50 Hz (Lead /Lag)	0.01 PF to 1 PF	0.003PF	Using Multifunction calibrator Fluke 5522A By direct method 50Hz
48	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Resistance	100 Mohm to 1000 Mohm	0.062% to 1.82%	Using Multifunction Calibrator Fluke 5522A by Direct method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 57 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
Site Facility					
49	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Voltage (10 k Hz to 50 kHz)	3 mV to 329 V	0.39% to 0.046%	Using Multifunction calibrator Flue 5522A By direct method
50	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Voltage (100kHz to 500kHz)	3 mV to 3 V	2.85% to 0.32%	Using Multifunction calibrator Flue 5522A By direct method
51	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Voltage (10Hz to 45Hz)	3 mV to 32 V	0.32% to 0.037%	Using Multifunction calibrator Flue 5522A By direct method
52	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Voltage (20kHz to 50kHz)	100 mV to 300 V	0.058% to 0.04%	Using Multifunction calibrator Flue 5522A By direct method
53	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Voltage (20kHz to 50kHz)	3 mV to 100 mV	0.42% to 0.058%	Using Multifunction calibrator Flue 5522A By direct method
54	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Voltage (45 Hz to 10kHz)	1 mV to 100 mV	0.96% to 0.028%	Using Multifunction calibrator Flue 5522A By direct method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 58 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
Site Facility					
55	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Voltage (45 Hz to 10kHz)	100 mV to 1000 V	0.028% to 0.037%	Using Multifunction calibrator Fluke 5522A By direct method
56	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Voltage (50 kHz to 100 kHz)	3 mV to 329 V	0.90% to 0.27%	Using Multifunction calibrator Fluke 5522A By direct method
57	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	DC Current	1 A to 10 A	0.08% to 0.19%	Using Fluke 8846A 6½ DMM Direct method
58	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	DC Current	10 µA to 1 A	0.35% to 0.081%	Using Fluke 8846A 6½ DMM with by Direct method
59	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	DC Current	10 A to 20 A	0.19% to 0.37%	Using Fluke 8846A 6½ DMM with 30A Shunt Direct method
60	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	DC Current	20 A to 750 A	0.36% to 0.75%	Using Std shunt with Fluke 8846A 6½ DMM Direct method
61	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	DC High Voltage	1 kV to 10 kV	3.20%	Using HV probe with DMM by comparison Method
62	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	DC High Voltage	10 kV to 40 kV	3.20%	Using HV probe with DMM by comparison Method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 59 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
Site Facility					
63	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	DC Voltage	0.1 mV to 1 mV	4.25% to 0.43%	Using Fluke 8846A 6½ DMM Direct method
64	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	DC Voltage	1 mV to 1000 V	0.42% to 0.006%	Using Fluke 8846A 6½ DMM by Direct method
65	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	Resistance	0.1 ohm to 1 ohm	3.48% to 0.36%	Using Fluke 8846A 6½ DMM Direct method
66	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	Resistance	1 Mohm to 10 Mohm	0.013% to 0.05%	Using Fluke 8846A 6½ DMM Direct method
67	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	Resistance	1 ohm to 1 Mohm	0.36% to 0.01%	Using Fluke 8846A 6½ DMM Direct method
68	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	Resistance	10 Mohm to 100 Mohm	0.05% to 0.93%	Using Fluke 8846A 6½ DMM Direct method
69	ELECTRO-TECHNICAL- DIRECT CURRENT (Measure)	Resistance	100 Mohm to 1000 Mohm	0.93% to 2.34%	Using Fluke 8846A 6½ DMM Direct method
70	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	DC Current	1 µA to 10 µA	2.45% to 0.25%	Using Multifunction calibrator Fluke 5522A By direct method
71	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	DC current	1 A to 10 A	0.028% to 0.064%	Using Multifunction calibrator Fluke 5522A By direct method
72	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	DC current	10 µA to 100 mA	0.25% to 0.016%	Using Multifunction calibrator Fluke 5522A By direct method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 60 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
Site Facility					
73	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	DC current	10 A to 20 A	0.064% to 0.12%	Using Multifunction calibrator Flue 5522A By direct method
74	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	DC current	100 mA to 1 A	0.016% to 0.028%	Using Multifunction calibrator Flue 5522A By direct method
75	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	DC current	20 A to 1000 A	0.82% to 0.71%	Using Multifunction calibrator Flue 5522A with Current Coil By Direct Method
76	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	DC Voltage	0.1 mV to 1 mV	1.5% to 0.24%	Using Multifunction calibrator Flue 5522A By direct method
77	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	DC Voltage	1 mV to 100 mV	0.24% to 0.0039%	Using Multifunction calibrator Flue 5522A By direct method
78	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	DC Voltage	10 V to 1000 V	0.0019% to 0.0025%	Using Multifunction calibrator Flue 5522A By direct method
79	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	DC Voltage	100 mV to 10 V	0.0039% to 0.0018%	Using Multifunction calibrator Flue 5522A By direct method
80	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	Resistance	0.1 Mohm to 10 Mohm	0.82% to 0.58%	Using High stability Decade Megaohm Box Vaiseshika 8400HV by direct Method
81	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	Resistance	1 Gohm to 100 Gohm	1.38 % to 2.32 %	Using High stability Decade Megaohm Box 8400 HV by direct Method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 61 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
Site Facility					
82	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	Resistance	1 Mohm to 100 Mohm	0.0043% to 0.062%	Using Multifunction Calibrator Fluke 5522A by Direct method
83	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	Resistance	1 ohm to 10 kohm	0.061%	Using High Precision Decade Resistance Box Vaiseshika 7400 by direct Method
84	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	Resistance	10 Mohm to 1 Gohm	0.58% to 1.38%	Using High stability Decade Megaohm Box Vaiseshika 8400 by direct Method
85	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	Resistance	100 Gohm to 1 Tohm	2.32% to 3.50%	Using High stability Decade Megaohm Box 8400 HV by direct Method
86	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	Resistance	100 mohm to 1 Mohm	1.52 % to 0.0043 %	Using Multifunction Calibrator Fluke 5522A by Direct method
87	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	Resistance (4wire)	10 mohm to 100 mohm	5.77% to 1.52%	Using Multifunction calibrator Fluke 5522A By Direct Method
88	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	Resistance (Micro-ohm Meters)	1 mohm @ 31.6A	0.12 %	Using Micro ohm calibrator Vaiseshika 9409-CAL by direct Method
89	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	Resistance (Micro-ohm Meters)	10 mohm @10A	0.06 %	Using Micro ohm calibrator Vaiseshika 9409-CAL by direct Method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 62 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
Site Facility					
90	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	Resistance (Micro-ohm Meters)	10μohm to 100μohm @ 200 A	4.33 % to 1.07 %	Using Micro ohm calibrator Vaiseshika 9409-CAL by direct Method
91	ELECTRO-TECHNICAL- DIRECT CURRENT (Source)	Resistance (Micro-ohm Meters)	100 mohm @3.16A	0.03 %	Using Micro ohm calibrator Vaiseshika 9409-CAL by direct Method
92	ELECTRO-TECHNICAL- ELECTRICAL EQUIPMENT (Measure)	CT Burden (1A/5A)	1.25 VA to 50 VA	1.32% to 1.38%	Using Eltel Bridge AITTS-98 by Direct Method
93	ELECTRO-TECHNICAL- ELECTRICAL EQUIPMENT (Measure)	CT Phase Error	3000A /1A & 5 A to 6000A / 1A & 5 A	2.25min	Using Std CT 0.05 & Eltel Bridge(AITTS-98) byComparison Method
94	ELECTRO-TECHNICAL- ELECTRICAL EQUIPMENT (Measure)	CT Phase Error	5A/1A & 5 A to 3200A/1A & 5 A	1.04min	Using Std CT 0.05 & Eltel Bridge AITTS-98 Direct/comparison Method
95	ELECTRO-TECHNICAL- ELECTRICAL EQUIPMENT (Measure)	CT Ratio Error	3000A / 1A & 5 A to 6000A / 1A & 5 A	0.084%	Using Std CT 0.05 & Eltel Bridge AITTS-98 Direct/comparison Method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 63 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
Site Facility					
96	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Measure)	CT Ratio Error	5A / 1A & 5 A to 3200A / 1A & 5 A	0.059%	Using Std CT 0.05 & Eltel Bridge AITTS-98 Direct/comparison Method
97	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Measure)	PT Burden (63.5 / 110V)	2.5 VA to 200 VA	1.25%	Using Eltel Bridge AITTS-98 by Direct Method
98	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Source)	Oscilloscope Bandwidth 3dB	50 kHz to 1 GHz	2.43% to 6.03%	Using Multifunction Calibrator Fluke 5522A Direct Method
99	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Source)	Oscilloscope DC Amplitude 1Mohm output	8 mV to 100 V	0.64% to 0.060%	Using Multifunction Calibrator Fluke 5522A Direct Method
100	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Source)	Oscilloscope Time Base	1 nS to 5 S	0.076% to 0.58%	Using Multifunction Calibrator Fluke 5522A Direct Method
101	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Source)	Oscilloscope AC Amplitude 1 Mohm output @1kHz (Vpp)	8 mV to 100 V	0.98% to 0.12%	Using Multifunction Calibrator Fluke 5522A Direct Method
102	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Indicator/Recorder/Controller/ScannerB -Type	600 °C to 1820 °C	1.41°C to 0.61°C	Using Fluke 8846A 6½ DMM using e for ITS-90 scale mV to °C conversion Method by Direct Method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 64 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
Site Facility					
103	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Indicator/Recorder/Controller/ScannerC -Type	10 °C to 1820 °C	0.30°C to 0.31°C	Using Fluke 8846A 6½ DMM using e for ITS-90 scale mV to °C conversion Method by Direct Method
104	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Indicator/Recorder/Controller/ScannerE -Type	-250 °C to 1000 °C	0.094°C to 0.10°C	Using Fluke 8846A 6½ DMM using e for ITS-90 scale mV to °C conversion Method by Direct Method
105	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Indicator/Recorder/Controller/ScannerJ -Type	-210 °C to 1200 °C	0.096°C to 0.12°C	Using Fluke 8846A 6½ DMM using e for ITS-90 scale mV to °C conversion Method by Direct Method
106	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Indicator/Recorder/Controller/ScannerK-Type	-200 °C to 1372 °C	0.15°C to 0.19°C	Using Fluke 8846A 6½ DMM using e for ITS-90 scale mV to °C conversion Method by Direct Method
107	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Indicator/Recorder/Controller/ScannerL -Type	-200 °C to 900 °C	0.09°C to 0.11°C	Using Fluke 8846A 6½ DMM using e for ITS-90 scale mV to °C conversion Method by Direct Method
108	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Indicator/Recorder/Controller/ScannerN -Type	-200 °C to 1300 °C	0.19°C to 0.17°C	Using Fluke 8846A 6½ DMM using e for ITS-90 scale mV to °C conversion Method by Direct Method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 65 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
Site Facility					
109	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Indicator/Recorder/Controller/ScannerR -Type	-40 °C to 1767 °C	0.88°C to 0.42°C	Using Fluke 8846A 6½ DMM using e for ITS-90 scale mV to °C conversion Method by Direct Method
110	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Indicator/Recorder/Controller/ScannerRTD (PT-100)	-200 °C to 850 °C	0.10°C to 0.12°C	Using Fluke 8846A 6½ DMM using e for ITS-90 scale ohm to °C conversion Method by Direct Method
111	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Indicator/Recorder/Controller/ScannerS -Type	-40 °C to 1767 °C	0.86°C to 0.46°C	Using Fluke 8846A 6½ DMM using e for ITS-90 scale mV to °C conversion Method by Direct Method
112	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Indicator/Recorder/Controller/ScannerT -Type	-250 °C to 400 °C	0.15°C to 0.10°C	Using Fluke 8846A 6½ DMM using e for ITS-90 scale mV to °C conversion Method by Direct Method
113	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Indicator/Recorder/Controller/ScannerU -Type	-200 °C to 600 °C	0.09°C to 0.11°C	Using Fluke 8846A 6½ DMM using e for ITS-90 scale mV to °C conversion Method by Direct Method
114	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	B Type	600 °C to 1820 °C	0.21°C	Using Multifunction Calibrator Fluke 5522A Direct /using ITS-90 scale for mV to °C conversion Method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 66 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
Site Facility					
115	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	C Type	0 °C to 2316 °C	0.07°C to 0.09°C	Using Multifunction Calibrator Fluke 5522A Direct /using ITS-90 scale for mV to °C conversion Method
116	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	E Type	-250 °C to 1000 °C	0.33°C to 0.017°C	Using Multifunction Calibrator Fluke 5522A Direct /using ITS-90 scale for mV to °C conversion Method
117	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	J Type	-210 °C to 1200 °C	0.10°C to 0.09°C	Using Multifunction Calibrator Fluke 5522A Direct /using ITS-90 scale for mV to °C conversion Method
118	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	K -Type	-200 °C to 1372 °C	0.13°C to 0.09°C	Using Multifunction Calibrator Fluke 5522A Direct /using ITS-90 scale for mV to °C conversion Method
119	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	L Type	-200 °C to 900 °C	0.06°C	Using Multifunction Calibrator Fluke 5522A Direct /using ITS-90 scale for mV to °C conversion Method
120	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	N Type	-200 °C to 1300 °C	0.14°C to 0.09°C	Using Multifunction Calibrator Fluke 5522A Direct /using ITS-90 scale for mV to °C conversion Method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 67 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
Site Facility					
121	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	R Type	0 °C to 1767 °C	0.31°C to 0.15°C	Using Multifunction Calibrator Fluke 5522A Direct /using ITS-90 scale for mV to °C conversion Method
122	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	RTD (Pt 100)(2 /4 wire)	-200 °C to 800 °C	0.06°C to 0.27°C	Using Multifunction Calibrator Fluke 5522A Direct /using ITS-90 scale for ohm to °C conversion Method
123	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	S Type	0 °C to 1767 °C	0.30°C to 0.17°C	Using Multifunction Calibrator Fluke 5522A Direct /using ITS-90 scale for mV to °C conversion Method
124	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	T Type	-250 °C to 400 °C	0.22°C to 0.07°C	Using Multifunction Calibrator Fluke 5522A Direct /using ITS-90 scale for mV to °C conversion Method
125	ELECTRO-TECHNICAL- TIME & FREQUENCY (Measure)	Frequency	1 Hz to 1 GHz	0.0082% to 0.006%	Using High resolution counter By direct method
126	ELECTRO-TECHNICAL- TIME & FREQUENCY (Measure)	Time	30 min to 24 Hrs	2.32S to 6S	Using Digital Stop Watch By comparison Method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 68 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
Site Facility					
127	ELECTRO-TECHNICAL- TIME & FREQUENCY (Measure)	Time	6 S to 30 min	0.2S to 2.32S	Using Digital Stop Watch By comparison Method
128	ELECTRO-TECHNICAL- TIME & FREQUENCY (Source)	Frequency	1 Hz to 10 Hz	0.0059% to 0.0008%	Using Multifunction Calibrator Fluke 5522A by Direct method
129	ELECTRO-TECHNICAL- TIME & FREQUENCY (Source)	Frequency	10 Hz to 1 MHz	0.0007% to 0.003%	Using Multifunction calibrator Fluke 5522A By direct method
130	ELECTRO-TECHNICAL- TIME & FREQUENCY (Source)	Frequency	100 kHz to 1 GHz	0.007% to 0.0003 %	Using Signal generator HP By Direct method
131	ELECTRO-TECHNICAL- TIME & FREQUENCY (Source)	Time & Period	2 nS to 5 S	1.17% to 0.001%	Using Multifunction calibrator Fluke 5522A By direct method
132	MECHANICAL-ACCELERATION AND SPEED	RPM Indicator / Centrifuge	100 to 5000 rpm	1 rpm to 3.5rpm	Using Digital Tachometer By comparison Method
133	MECHANICAL-ACCELERATION AND SPEED	RPM Indicator / Centrifuge	5000 to 10000 RPM	3.5 rpm to 33.0rpm	Using Digital Tachometer By comparison Method
134	MECHANICAL-ACCELERATION AND SPEED	Tachometer , Calibrator (Non contact Type)	100 to 1000 rpm	0.2rpm to 1.9rpm	Using Digital Tachometer & RPM Calibrator By comparison Method
135	MECHANICAL-ACCELERATION AND SPEED	Tachometer , Calibrator (Non contact Type)	1000 rpm to 10000 rpm	1.9rpm to 14.8rpm	Using Digital Tachometer & RPM Calibrator By comparison Method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 69 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
Site Facility					
136	MECHANICAL-ACCELERATION AND SPEED	Tachometer , Calibrator (Non contact Type)	10000 rpm to 50000 rpm	14.74 rpm to 32.72 rpm	Using Digital Tachometer & RPM Calibrator By comparison Method
137	MECHANICAL-ACCELERATION AND SPEED	Tachometer , Calibrator (Non contact Type)	50000 to 100000 rpm	33rpm to 61.0rpm	Using Digital Tachometer & RPM Calibrator By comparison Method
138	MECHANICAL-ACCELERATION AND SPEED	Tachometer , Calibrator (contact Type)	100 rpm to 1000 rpm	1.0rpm to 2.2rpm	Using Digital Tachometer & RPM Calibrator By comparison Method
139	MECHANICAL-ACCELERATION AND SPEED	Tachometer , Calibrator (contact Type)	1000 rpm to 5000 rpm	2.2rpm to 4.0rpm	Using Digital Tachometer & RPM Calibrator By comparison Method
140	MECHANICAL-ACOUSTICS	Sound Level Meter	94 dBA & 114 dBA	2.4dBA	using Sound level Calibrator by comparison
141	MECHANICAL-PRESSURE INDICATING DEVICES	Absolute Pressure Gauge,/Digital / Analog / Pressure Gauge, Pressure Transmitter, Pressure Switch, Barometer	0.1 to 35(Abs) bar	0.022bar	Digital Pressure Indicator with Pneumatic Pump DKD -R-6-1
142	MECHANICAL-PRESSURE INDICATING DEVICES	Absolute Pressure Gauge,/Digital / Analog / Pressure Gauge,Pressure Transmitter, Pressure Switch, Barometer	0.1 to 7 (Abs) bar	0.008bar	Digital Pressure Indicator with Pneumatic Pump DKD -R-6-1



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 70 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
Site Facility					
143	MECHANICAL- PRESSURE INDICATING DEVICES	Absolute Pressure Gauge,/Digital / Analog / Pressure Gauge,Vacuum Gauge , Vacuum Indicator, Vacuum Transmitter, Pressure Transmitter, Pressure Switch, Barometer	60 to 110 (Abs) kPa	0.266kPa	Digital Pressure Indicator with Pneumatic Pump DKD -R-6-1
144	MECHANICAL- PRESSURE INDICATING DEVICES	Digital / Analog Pressure Gauge, Indicator ,Manometer, Magnehelic Gauge,Differential Pressure Transmitter,Pressure Switches	0 to 200(g) mbar	0.27mbar	Digital Pressure Indicator with Pneumatic Pump DKD -R-6-1
145	MECHANICAL- PRESSURE INDICATING DEVICES	Digital / Analog Pressure Gauge, Indicator ,Manometer, Magnehelic Gauge,Differential Pressure Transmitter,Pressure Switches	-12.5 mbar to 12.5(g) mbar	0.037mbar	Digital Pressure Indicator with Pneumatic Pump DKD -R-6-1



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 71 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
Site Facility					
146	MECHANICAL- PRESSURE INDICATING DEVICES	Digital / Analog Pressure Gauge, Indicator ,Manometer, Magnehelic Gauge,Differential Pressure Transmitter,Pressure Switches	-2.5 mbar to 2.5 (g) mbar	0.014mbar	Digital Pressure Indicator with Pneumatic Pump DKD -R-6-1
147	MECHANICAL- PRESSURE INDICATING DEVICES	Digital / Analog Pressure Gauge, Indicator ,Manometer, Magnehelic Gauge,Differential Pressure Transmitter,Pressure Switches	-25 to 25(g) mbar	0.029mbar	Digital Pressure Indicator with Pneumatic Pump DKD -R-6-1
148	MECHANICAL- PRESSURE INDICATING DEVICES	Digital / Analog / Vacuum Gauge, Indicator, Vacuum Transmitter	-0.93 bar to 0 bar	0.007bar	Using Digital Vacuum gauge with Pneumatic Pump DKD-R-6-2
149	MECHANICAL- PRESSURE INDICATING DEVICES	Pressure Hydraulic Digital / Analog Pressure Gauge, Indicator, Pressure Transmitter,Pressure Switches	0 to 70 bar	0.020bar	Digital Pressure Indicator with Pneumatic Pump DKD -R-6-1



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 72 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
Site Facility					
150	MECHANICAL- PRESSURE INDICATING DEVICES	Pressure HydraulicDigital / Analog Pressure Gauge, Indicator, Pressure Transmitter,Pressure Switches	0 to 350 bar	0.15bar	Digital Pressure Indicator with Hydraulic Pump DKD -R-6-1
151	MECHANICAL- PRESSURE INDICATING DEVICES	Pressure HydraulicDigital / Analog Pressure Gauge, Indicator, Pressure Transmitter,Pressure Switches	0 bar to 700 bar	0.19bar	Digital Pressure Indicator with Hydraulic Pump DKD -R-6-1
152	MECHANICAL- PRESSURE INDICATING DEVICES	Pressure Pneumatic Digital / Analog Pressure Gauge, Indicator, Pressure Transmitter,Pressure Switches	0 to 2 bar	0.001 bar	Digital Pressure Indicator with Pneumatic Pump DKD -R-6-1
153	MECHANICAL- PRESSURE INDICATING DEVICES	Pressure Pneumatic Digital / Analog Pressure Gauge, Indicator, Pressure Transmitter,Pressure Switches	0 to 20 bar	0.008bar	Digital Pressure Indicator with Pneumatic Pump DKD -R-6-1
154	MECHANICAL- PRESSURE INDICATING DEVICES	Pressure Pneumatic Digital / Analog Pressure Gauge, Indicator, Pressure Transmitter,Pressure Switches	0 bar to 35 bar	0.011bar	Digital Pressure Indicator with Pneumatic Pump DKD -R-6-1



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 73 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
Site Facility					
155	MECHANICAL- WEIGHING SCALE AND BALANCE	Weighing Machine Readability d = 1 mg	0 kg to 20 kg	10mg	Using weights of accuracy class E1 & E 2 procedure based on OIML R-76 (2006)
156	MECHANICAL- WEIGHING SCALE AND BALANCE	Weighing Machine Readability d =0.1 mg	Upto 2300 g	0.78mg	Using weights of accuracy class E1 & E 2 procedure based on OIML R-76 (2006)
157	MECHANICAL- WEIGHING SCALE AND BALANCE	Weighing Machine Readability d =0.1g	0 kg to 35 kg	0.26g	Using weights of accuracy class E 2 procedure based on OIML R-76 (2006)
158	MECHANICAL- WEIGHING SCALE AND BALANCE	Weighing Machine Readability d =10 g	Upto 200 kg	11g	Using Weights of Accuracy class F1 , Procedure based on OIML R-76(2006)
159	MECHANICAL- WEIGHING SCALE AND BALANCE	Weighing Machine Readability d=0.001mg	0 g to 5 g	0.005mg	Using weights of accuracy class E1 procedure based on OIML R-76 (2006)
160	MECHANICAL- WEIGHING SCALE AND BALANCE	Weighing Machine Readability d=0.01mg	Upto 200 g	0.06mg	Using weights of accuracy class E1 procedure based on OIML R-76 (2006)
161	MECHANICAL- WEIGHING SCALE AND BALANCE	Weighing Machine Readability d=50 g	Upto 400 kg	43g	Using Weights of Accuracy class F1 , Procedure based on OIML R-76(2006)



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 74 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
Site Facility					
162	THERMAL- SPECIFIC HEAT & HUMIDITY	Humidity calibrator, Generator, environmental chamber at single point	10 to 90%RH @25°C	0.48%RH	Using RH & Temp. Probe with Indicator By Comparison Method
163	THERMAL- SPECIFIC HEAT & HUMIDITY	Humidity sensor with indicator, Transmitter with sensor, Hygrometers, Humidity data loggers with sensors	10% to 90% @25°C	0.52%RH	Using RH Generator and RH & Temp. Probe with Indicator By Comparison Method
164	THERMAL- SPECIFIC HEAT & HUMIDITY	Temperature Humidity sensor with indicator, Transmitter with sensor, ThermoHygrometers, Temperature Humidity data loggers with sensors	10 °C to 50°C @50%RH	0.35°C	Using RH & Temp. Probe with Indicator By Comparison Method
165	THERMAL- TEMPERATURE	Environmental Chamber, Cold Rooms, Storage Room, Deep Freezer, Dry Well, Cooling Cabinet	-80 °C to 300 °C	1.85°C	Using RTD PT-100 Sensors with recorder (Min Nine) Multi position Calibration
166	THERMAL- TEMPERATURE	RTD, /PRT/TC Sensor with & Without Indicators , Digital thermometer, Temp Transmitter with Sensor, Recorder with sensor , Thermistor with temp Indicator.	(-)80 °C to 0 °C	0.046°C	PRT with Precision Scanner & Liquid Bath By Comparison Method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 75 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
Site Facility					
167	THERMAL-TEMPERATURE	RTD,/PRT/TC Sensor with & Without Indicators , Digital thermometer,Temp Transmitter with Sensor, Recorder with sensor ,Thermistor with temp Indicator.	0 °C to 100 °C	0.049°C	Using PRT with Precision Scanner & Dry Bath By Comparison Method
168	THERMAL-TEMPERATURE	RTD,/PRT/TC Sensor with & Without Indicators , Digital thermometer,Temp Transmitter with Sensor, Recorder with sensor ,Thermistor with temp Indicator.	0 °C to 140 °C	0.07°C	PRT with Precision Scanner & Dry Bath By Comparison Method
169	THERMAL-TEMPERATURE	RTD,/PRT/TC Sensor with & Without Indicators , Digital thermometer,Temp Transmitter with Sensor, Recorder with sensor ,Thermistor with temp Indicator.	100 °C to 200 °C	0.062°C	PRT with Precision Scanner & Liquid Bath By Comparison Method
170	THERMAL-TEMPERATURE	RTD,/PRT/TC Sensor with & Without Indicators , Digital thermometer,Temp Transmitter with Sensor, Recorder with sensor ,Thermistor with temp Indicator.	140 °C to 660 °C	0.24°C	Using PRT with Precision Scanner & Dry Bath By Comparison Method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 76 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
Site Facility					
171	THERMAL-TEMPERATURE	RTD,/PRT/TC Sensor with & Without Indicators , Digital thermometer,Temp Transmitter with Sensor, Recorder with sensor ,Thermistor with temp Indicator.	200 °C to 250 °C	0.07°C	PRT with Precision Scanner & Liquid Bath By Comparison Method
172	THERMAL-TEMPERATURE	RTD,/PRT/TC Sensor with & Without Indicators , Digital thermometer,Temp Transmitter with Sensor, Recorder with sensor ,Thermistor with temp Indicator.	-95 °C to 0 °C	0.07°C	Using PRT with Precision Scanner & Dry Bath By Comparison Method
173	THERMAL-TEMPERATURE	Temp.Indicator with sensor of Deep Freezer,Oven,Chamber ,Incubators,Oil Bath,Dry well cooling Cabinet,	-95 °C to 0 °C	0.017°C to 0.042°C	Using PRT with Precision Scanner single position calibration
174	THERMAL-TEMPERATURE	Temp.Indicator with sensor of Deep Freezer,Oven,Chamber ,Incubators,Oil Bath,Dry well cooling Cabinet, Muffle Furnace .	0 °C to 660 °C	0.042°C to 0.052°C	Using PRT with Precision Scanner single position calibration



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 77 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
Site Facility					
175	THERMAL-TEMPERATURE	Temp.Indicator with sensor of Deep Freezer,Oven,Chamber ,Incubators,Oil Bath,Dry well cooling Cabinet, Muffle Furnace .	1100 °C to 1200 °C	2.583°C	Using R-Type Thermocouple with Precision Scanner single position calibration
176	THERMAL-TEMPERATURE	Temp.Indicator with sensor of Deep Freezer,Oven,Chamber ,Incubators,Oil Bath,Dry well cooling Cabinet, Muffle Furnace .	600 °C to 800 °C	1.58°C	Using R-Type Thermocouple with Precision Scanner single position calibration
177	THERMAL-TEMPERATURE	Temp.Indicator with sensor of Deep Freezer,Oven,Chamber ,Incubators,Oil Bath,Dry well cooling Cabinet, Muffle Furnace .	800 °C to 1100 °C	1.8°C	Using R-Type Thermocouple with Precision Scanner single position calibration
178	THERMAL-TEMPERATURE	Thermocouple with & Without Indicators , Digital thermometer,Temp Transmitter with Sensor	1100 °C to 1200 °C	1.892°C to 2.583°C	Using R-Type Thermocouple with Precision scanner & furnace By Comparison Method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name AUTOCAL SOLUTIONS PVT LTD, B-80, MIDC, SATPUR, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2052 Page No. : 78 / 78

Validity 26/04/2019 to 25/04/2021 Last Amended on -

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(\pm)	Remarks
Site Facility					
179	THERMAL-TEMPERATURE	Thermocouple with & Without Indicators , Digital thermometer, Temp Transmitter with Sensor	800 °C to 1100 °C	1.93°C	Using R-Type Thermocouple with Precision scanner & furnace By Comparison Method
180	THERMAL-TEMPERATURE	Thermocouple with & Without Indicators , Digital thermometer, Temp Transmitter with Sensor , Thermistor with temp Indicator.	600 °C to 800 °C	1.77°C	Using R-Type Thermocouple with Precision scanner & furnace By Comparison Method

* CMCs represent expanded uncertainties expressed at approximately the 95% level of confidence, using a coverage factor of $k = 2$.